



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NC SL Z540-1-1994 & ANSI/NC SL Z540.3-2006

MOREHOUSE INSTRUMENT CO., INC.
1742 Sixth Avenue
York, PA 17403-2675
Harry E. Zumbrun Phone: 717 843 0081

CALIBRATION

Valid to: April 30, 2020

Certificate Number: 1398.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
DC Voltage – Electrical Calibration of Load Indicators	(0 to 4.4) mV/V	0.00005 mV/V	Load cell simulator

II. Mechanical

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Force – Dead Weight Primary	(5 to 105) gf	0.003 %	Force Calibration including ASTM E74 Class A and AA, ISO 376 Class 00, 0.5, 1 and 2 Forces can be applied incrementally and decrementally through 120 000 lbf thus permitting the determination of hysteresis errors
Standards: Tension and Compression	(0.1 to 10) lbf [(0.44 to 44) N]	0.0025 %	
	(10 to 100) lbf [(44 to 444) N]	0.0016 %	
	(100 to 12 000) lbf [(444 to 53 378) N]	0.0016 %	
	(12 000 to 120 000) lbf [(53 378 to 533 786) N]	0.0016 %	

Parameter/Equipment	Range	CMC ^{2,3,4} (±)	Comments
Force (cont) – N.I.S.T Calibrated Transfer/Secondary Standards: Tension and Compression	(120 000 to 1 000 000) lbf [(533 to 4448) kN]	1.5E-05 x F + 11 [(11.5 through 26.17) lbf] [(51.2 through 116.4) N]	Force Calibration including ASTM E74 Class A, ISO 376 Class 0, 0.5, 1 and 2 Forces can be applied incrementally and decrementally through 1 000 000 lbf thus permitting the determination of hysteresis errors
N.I.S.T Calibrated Transfer/Secondary Standards: Compression	(1 000 000 to 2 250 000) lbf [(4.4 to 10) MN]	4.22E-05 x F + 52. [(94.7 through 147.6) lbf] [(421.2 through 655.9) N]	Forces can be applied incrementally only from 1 000 000 through 2 250 000 lbf
Tension	(1 000 000 to 1 125 000) lbf [(4.4 to 5) MN]	4.22E-05 x F + 52.5 [(94.7 through 145.3) lbf] [(421.2 through 441.71) N]	
Aircraft Scales/Truck Scales (Portable) ⁵	(0 to 60 000) lbf	0.0016 %	Force



Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Torque –			
Dead Weight Primary Standards	(0.37 to 73.75) lbf·ft; (0.5 to 100) N·m	0.005 %	Primary torque standard, ASTM E2428 and other methods
Clockwise & Counter-clockwise	(14.75 to 1475) lbf·ft; (20 to 2000) N·m	0.003 %	

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC Uncertainty due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, percentages are to read as percent of the indicated value, unless otherwise noted.

⁴ In the statement of CMC, F = Applied force in lbf.

⁵ The CMC for this Parameter/Equipment applies for performance verification of the “best existing” device under test and not for the assignment of reference values, and therefore certain characteristics of the “best existing” device under test (e.g. resolution) are not included in this CMC estimate.



Accredited Laboratory

A2LA has accredited

MOREHOUSE INSTRUMENT COMPANY., INC.

York, PA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSLI Z540-1-1994 and the requirements of ANSI/NCSLI Z540.3-2006 and R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 9th day of May 2018.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 1398.01
Valid to April, 30, 2020
Revised August 3, 2018

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.