

Converting an mV/V load cell signal into Engineering Units: Why this is the most accurate and cost-effective way to use a calibration curve

Understanding what mV/V is and how does it relate to load cells

Most bridge-based sensors typically specify a rated output Sensitivity (R.O.) shown in figure 1 below. This Rated Output is typically found under Electrical specifications. It is usually in mV/V, where mV/V is the ratio of the output voltage to the excitation voltage required for the sensor to work. Most load cells are strain gauge-based sensors that provide a voltage output that is proportional to the excitation voltage. Many feature four strain gauges in a Wheatstone bridge configuration. When force is applied, the relative change in resistance is what is measured by the indicator. The load cell signal is converted to a visual or numeric value by a “digital indicator.” When there is no load on the cell, the two signal lines are at equal voltage. As a load is applied to the cell, the voltage on one signal line increases very slightly, and the voltage on the other signal line decreases very slightly. The difference in voltage between the two signals is read by the indicator. Recording these reading in mV/V is often the most accurate method for measurement. The reason it is the most accurate method is many meters on the market can handle ratiometric measurements. They can measure the input in mV and divide that measurement by the actual voltage being supplied. For instance, we could have an mV measurement of 40.1235 mV and an excitation measurement of 9.9998 V. When displaying in mV/V one would have 4.01243 mV/V. Many meters that do not handle ratiometric measurements, they have some internal counts that get programmed at the time of calibration. These meters still read the change in resistance; they require programing or points to be entered that correspond to force values.



Specifications	Model - Capacity (lbf / kN)				
	300-2K / 1-10	5K-10K / 20-50	25K-50K / 100-250	60K / 300	100K / 500
Accuracy					
Static Error Band, % R.O.	± 0.02	± 0.03	± 0.04	± 0.04	± 0.04
Non-Linearity, % R.O.	± 0.03	± 0.03	± 0.03	± 0.03	± 0.03
Hysteresis, % R.O.	± 0.02	± 0.04	± 0.04	± 0.04	± 0.04
Non-Repeatability, % R.O.	± 0.01	± 0.01	± 0.01	± 0.01	± 0.01
Creep, % Rdg / 20 Min.	± 0.03	± 0.03	± 0.03	± 0.03	± 0.03
Off-Center Load Sensitivity, %/in	± 0.10	± 0.10	± 0.10	± 0.10	± 0.10
Side Load Sensitivity, %	± 0.10	± 0.10	± 0.10	± 0.10	± 0.10
Zero Balance, % R.O.	± 1.0	± 1.0	± 1.0	± 1.0	± 1.0
Temperature					
Range, Compensated, °F	+15 to +115	+15 to +115	+15 to +115	+15 to +115	+15 to +115
Range, Operating, °F	-65 to +200	-65 to +200	-65 to +200	-65 to +200	-65 to +200
Sensitivity Effect, % Rdg / 100°F	0.08	0.08	0.08	0.08	0.08
Zero Effect, % R.O. / 100°F	0.08	0.08	0.08	0.08	0.08
Electrical					
Recommended Excitation, VDC	10	10	10	10	10
Input Resistance, Ω	350 +40/-3.5	350 +40/-3.5	350 +40/-3.5	350 +40/-3.5	350 +40/-3.5
Output Resistance, Ω	350 ± 3.5	350 ± 3.5	350 ± 3.5	350 ± 3.5	350 ± 3.5
Sensitivity (R.O.), mV/V, Nominal	2	4	4	4	4
Insulation Bridge/Case, MegΩ	5000 @ 50 VDC	5000 @ 50 VDC	5000 @ 50 VDC	5000 @ 50 VDC	5000 @ 50 VDC
Mechanical					
Safe Overload, % R.O.	150	150	150	150	150
Weight, lbs	1.0	2.9	9.1	11.2	23.5
Weight w/Base, lbs	2.5	6.5	21.5	26	52.5
Flexure Material	Aluminum	Steel	Steel	Steel	Steel

Figure 1 Morehouse Precision Shear Web Load Cell Specification Sheet

Programming a load cell system via span points

Most indicators will allow the end-user to span or capture data points. Several indicators offer many ways of programming points. Most of which are going to use some linear equation to display the non-programmed points along the curve or line.




Figure 2 Load Cell Curve Versus a Straight Line

My guess is almost everyone is familiar enough to remember Algebra and drawing a straight line between two points. One would typically find the slope of the line, which could predict other points along the line. The common formula of $y = mx + b$, where m designates the slope of the line, and where b is the y-intercept that is b is the second coordinate of a point where the line crosses the y-axis. The main issue with this approach when programming a load cell is that the meter and load cell are going to have some deviations from the straight line. A good indication of how much possible deviation is the Non-Linearity that is also found on the load cell specification sheet in figure 1. Non-Linearity is defined as the algebraic difference between OUTPUT at a specific load and the corresponding point on the straight line drawn between MINIMUM LOAD and MAXIMUM LOAD. There are other factors such as stability, thermal effects, creep recovery and return, and the loading conditions when the points are captured that will influence the bias of each point. The programming of these meters is going to follow a linear approach. Some will have a 2-pt span, some 5-pts, and some even more. They may try to draw a straight line through all the points, or they may try and segment several lines. In all cases, there will be additional bias created from this method as the force measuring system will always have some non-linear behavior.



Applied Force lbf	Actual Readings (mV/V)	Indicator with 2-pt adjustments		
		Programmed Points	Calculated Values 2 pt span	Error
200	0.08279		199.6	0.4
1000	0.41415	0.41415	998.6	1.4
2000	0.82851		1997.6	2.4
3000	1.24302		2997.0	3.0
4000	1.65767		3996.8	3.2
5000	2.07242		4996.8	3.2
6000	2.48726		5997.0	3.0
7000	2.90216		6997.4	2.6
8000	3.31709		7997.8	2.2
9000	3.73203		8998.3	1.7
10000	4.14696	4.14696	9998.7	1.3

Figure 3 Programming an Indicator with a 2-pt Span Calibration

Figure 3 above is an example of a Morehouse Calibration Shear Web Load Cell with a Non-Linearity specification of better than 0.05 % of full scale. In this example, the actual Non-Linearity is about 0.031 % using mV/V values and 0.032 % when using calculated values, which is well below the specification. However, one should never claim the device is accurate to 0.032 % as this is a short-term accuracy that was achieved under the ideal conditions. Often, an end-user will see the results above and make a claim that the system is accurate to a number such as 0.05 % and believe they are going to maintain it. However, the end-user must account for additional error sources such as stability/drift, reference standard uncertainty that was used to perform the calibration, resolution of the force measuring device, repeatability and reproducibility of the system, difference in loading conditions between the reference lab and how the system is being used, environmental conditions, and difference in adapters. All of which can drastically increase the overall accuracy specification. As a rule, accuracy is influenced by how the system is used, the frequency of calibration, the non-linearity of both the load cell and meter, as well as thermal characteristics. In addition, what the reference lab achieves is short term and does not include the stability of the system or adapters, which are often the most significant error sources. More information on adapters can be found [here](#).

Note: Several manufacturers claim specifications that use higher-order math equations for non-linearity to achieve unrealistic specifications. Especially, when programming a meter with these values. We generally find button or washer type load cells to have specifications that are very difficult to meet.

Figure 3 is an example of a 2-pt span calibration. Values are programmed at 1000 and 10,000 lbf. These values can often be entered into the meter or captured during setup with the force measuring system

under load. In the above example, one can see the instrument bias or error. Instrument Bias is defined in section 4.2 of JCGM 200:2012 as the average of replicate indications minus a reference quantity value. When we talk about bias, we are talking about the difference between the calculated values minus the applied force values. In the above example, the worst error is 3.2 lbf, which is around 0.08 % of applied force when 4000 lbf is applied.

Using Least Squares Method

Many indicators do not allow the end-user to enter anything other than span points. They do not allow the use of the “best-fit” or least-squares method. However, many indicators do have USB, IEEE, RS232, or other interfaces that will enable computers to read and communicate with the indicator. When software can communicate with an indicator, a method of regression analysis can be used, which often better characterizes the force measuring system. This method of regression analysis begins with a set of data points to be plotted on an x- and y-axis graph. The term “least squares” is used because it is the smallest sum of squares of errors. This method will contain a formula that is a bit more complex than a straight line. The formula often uses higher-order equations to minimize the error and best replicate the line. Figure 4 below shows a plot from the actual readings in mV/V and fit to a 3rd order equation. Instead of using the equation for a straight line ($y=mx+b$), we have a formula that uses x values that are raised to higher powers, such as $\text{Response (lb)} = A_0 + A_1F + A_2F^2$ Force (lbf) = where: F = Force (lbf) where: $R = A_0 = 0.0614$ $A_1 = 2415$ $A_2 = -1.4436$ $A_3 = 0.17379$. These are often called coefficients. They are often labeled as A0, A1, A2, A3. A0 would determine the point at which the equation crosses the Y-intercept, while the other coefficients determine the curve. Many force standards allow curve fitting of a 3rd degree and limit the maximum degree fit to a 5th degree. The most recognized legal metrology standards for using Coefficients are ASTM E74 and ISO 376. ASTM E74 Standard Practices for Calibration and Verification for Force-Measuring Instruments is primarily used in North America, while ISO 376 Metallic materials — Calibration of force-proving instruments used for the verification of uniaxial testing machines are used throughout much of Europe and the rest of the world. More information on these two standards can be found [here](#).




Figure 4 Graph of a 3rd Order Least Squares Fit

When the equation in figure 4 is used on the actual readings, the values calculated using the coefficients are very close to the applied force values. The bias or measurement error is around 0.1 lbf. I believe 0.1 lbf is less than the 3.2 lbf error as shown using a 2-pt span calibration 😊.

Applied Force lbf	Actual Readings (mV/V)	Using Coefficient Conversion	
		Calculated Values polynomial	Error
200	0.08279	199.9	0.1
1000	0.41415	999.9	0.1
2000	0.82851	1999.9	0.1
3000	1.24302	2999.9	0.1
4000	1.65767	3999.9	0.1
5000	2.07242	4999.9	0.1
6000	2.48726	5999.9	0.1
7000	2.90216	6999.9	0.1
8000	3.31709	7999.9	0.1
9000	3.73203	8999.9	0.1
10000	4.14696	9999.9	0.1

Figure 5 Bias or Measurement Error When Using Coefficients

The overall difference in the errors between these two methods is relatively high. Figure 6 below best summarizes these errors. One process produces an almost exact match, which is 0.001 % of full scale, while the other is 0.032 % of full scale. The worst point at 4,000 lbf has a difference of 3.06 lbf or 2413 %. The question is, what method do you think meets your needs? The process of using coefficients will often require additional software and a computer. The 2-pt adjustment will not. There are other considerations relating to calibration.




Figure 6 Difference Between 2-pt Span and Coefficients on the Same Load Cell

Calibration Differences

One of the more significant differences is with calibration. Any force measuring system is going to drift over time. The typical expectation of our customers is tweaking the units sent in for calibration, which is an attempt to minimize the bias. However, from the man who dropped his marbles, tweaking may not be right. If one is always adjusting the values or processes, it tends to become more out of control. It becomes more challenging to spot trends, which is an ISO/IEC 17025 requirement. Section 7.7.1 states, "The laboratory shall have a procedure for monitoring the validity of results. The resulting data shall be recorded in such a way that trends are detectable and, where practicable, statistical techniques shall be applied to review the results." With a span calibration that requires adjustments at every calibration interval, are trends truly detectable? When coefficients are used, the reference laboratory is merely reading the Actual Reading mV/V values at the time of each calibration. It is much easier to establish the baseline or monitoring the results based on units that are rarely adjusted. Note: Adjustments could happen if an indicator failed or a simulator is used to standardize the meter. Though that is another error source relating to the electrical side. If the indicator and load cell are paired and stay together as a system, this point is moot. It is highly recommended that one keeps their load cells and meter paired from one calibration to the next. When the reference laboratory reads and reports in mV/V using the least-squares method, ones "As Received" calibration becomes the same as the "As Returned." The end-user is given a new set of coefficients to use. The mV/V values are recorded and can be monitored. The

new coefficients will likely account for any drift that has happened and bring the force-measuring system back to having much lower bias than the span calibration.



Figure 7 Morehouse Load Cell System with Software

Morehouse software complies with ISO 376, ASTM E74, and E2428 requirements and eliminates the need to use load tables, excel reports, and other interpolation methods to ensure compliance with these standards. NCSLI RP-12 states in section 12.3 “The uncertainty in the value or bias, always increases with time since calibration”. When the drift occurs, the indicator needs to be reprogrammed, and most quality systems require an "As Received" calibration, then the indicator needs to be reprogrammed, and an "As Returned" calibration is performed. The actual level of work results in calibration costs that are much higher than they need to be. Morehouse developed our HADI and 4215 systems with software to avoid the excess costs as the coefficients used in the software are based on mV/V values, and the "As Received" and "As Returned" calibrations are the same with the end user only needing to update the coefficients in the software. The software allows for conversion from mV/V to lbf, kgf, kN, N and reduces the overall cost for the customer while meeting the quality requirements in ISO/IEC 17025:2017. Suppose additional software is a concern or problematic. In that case, we have a 4215 plus model that can store and use calibration coefficients that have a minimal error compared with traditional methods such as spanning multiple points.

Using mV/V Calibration Data and Entering Those Values into the Meter

COMPRESSION CALIBRATION DATA 3RD-ORDER FIT

FORCE APPLIED	MEASURED OUTPUT RUN 1 - 0°	MEASURED OUTPUT RUN 2 - 120°	MEASURED OUTPUT RUN 3 - 240°	FITTED CURVE	EXPANDED UNCERTAINTY	FORCE STANDARD
lbf	mV/V	mV/V	mV/V	mV/V	lbf	USED
100	-0.08336	-0.08337	-0.08342	-0.08339	0.0072	M-4644
500	-0.41671	-0.41674	-0.41678	-0.41674	0.0120	M-4644
1000	-0.83352	-0.83354	-0.83359	-0.83355	0.0210	M-4644
1500	-1.25046	-1.25046	-1.25050	-1.25046	0.0310	M-4644
2000	-1.66745	-1.66745	-1.66750	-1.66748	0.0410	M-4644
2500	-2.08457	-2.08456	-2.08460	-2.08458	0.0500	M-4644
3000	-2.50176	-2.50175	-2.50180	-2.50176	0.0600	M-4644
3500	-2.91902	-2.91901	-2.91905	-2.91901	0.0700	M-4644
4000	-3.33629	-3.33627	-3.33631	-3.33631	0.0800	M-4644
4500	-3.75365	-3.75364	-3.75367	-3.75365	0.0900	M-4644
5000	-4.17103	-4.17101	-4.17103	-4.17102	0.1000	M-4644

The Expanded Uncertainty is the aggregate uncertainty of the Morehouse measurement process, which includes the uncertainty of the reference standards used for calibration and the resolution of the unit under test. It is stated with a coverage factor of $k=2$, such that the confidence interval corresponds to approximately 95 %.

POLYNOMIAL EQUATIONS

The following polynomial equation, described in ASTM E74-18, has been fitted to the force and measured output values observed at calibration using the method of least squares

Response (mV/V) = $A_0 + A_1F + A_2F^2 + A_3F^3$
 where: F = Force (lbf)
 $A_0 = -5.868913E-05$
 $A_1 = -8.332379E-04$
 $A_2 = -2.666242E-10$
 $A_3 = 1.513019E-14$

Force (lbf) = $B_0 + B_1R + B_2R^2 + B_3R^3$
 where: R = Response (mV/V)
 $B_0 = -7.030104E-02$
 $B_1 = -1.200137E+03$
 $B_2 = -4.599537E-01$
 $B_3 = -3.135373E-02$

STANDARD DEVIATION	RESOLUTION	LOWER LIMIT FACTOR
mV/V	lbf	lbf
0.0000246	0.0120	0.0708

Figure 8 Calibration Report for a 5,000 lbf load cell



B Coefficients Additional Error				
mV/V	Predicted	Read on Meter	Difference	%
	LB S/N 12140			
-0.03999	47.92	47.92	0.00	0.000%
-0.07998	95.91	95.91	0.00	0.000%
-0.19995	239.88	239.88	0.00	0.000%
-0.39991	479.80	479.80	0.00	0.000%
-0.79979	959.51	959.51	0.00	0.000%
-1.19970	1439.13	1439.13	0.00	0.000%
-1.59962	1918.64	1918.64	0.00	0.000%
-1.99952	2398.04	2398.04	0.00	0.000%
-2.39942	2877.35	2877.34	0.01	0.000%
-3.19927	3835.81	3835.81	0.00	0.000%
-3.99901	4793.94	4793.94	0.00	0.000%
-4.39888	5272.96	5272.95	0.01	0.000%

Figure 9 5,000 lbf Morehouse Load Cell B Coefficient Error

Since this article was first published, we have done more testing on various scenarios using the formula for B coefficients embedded into a 4215 meter. We have developed an algorithm into the meter to display force values using the B coefficients in the above figure. When tested, the error from predicted was almost zero as there were some slight rounding errors as shown above. We know some people in the industry take the calibration reports and then enter mV/V into the meter. Thus, we decided to follow the same steps using a 5-pt and 2-pt calibration.

5 PT mV/V SPAN CALIBRATION				
mV/V	Predicted Force Values	Read on Meter LB S/N 12140	Difference	%
-0.03999	47.92	47.95	-0.03	-0.058%
-0.07998	95.91	95.91	0.00	0.004%
-0.19995	239.88	239.75	0.13	0.054%
-0.39991	479.80	479.54	0.26	0.055%
-0.79979	959.51	959.05	0.46	0.048%
-1.19970	1439.13	1438.59	0.54	0.037%
-1.59962	1918.64	1918.14	0.50	0.026%
-1.99952	2398.04	2397.51	0.53	0.022%
-2.39942	2877.35	2876.86	0.49	0.017%
-3.19927	3835.81	3835.33	0.48	0.013%
-3.99901	4793.94	4793.46	0.48	0.010%
-4.39888	5272.96	5272.52	0.44	0.008%
Programmed @ 1,2,3,4,5K				

Figure 10 5-PT mV/V Values Entered into the 4215 Meter



When we entered values programmed at 20 % increments and the corresponding mV/V values, the error on a device one expects to be better than 0.07 lbf (the ASTM LLF) is much higher at almost all test points. So the main issue here is if the end-user assumes they can do this and maintain the same uncertainty, they are mistaken.

2 PT mV/V SPAN CALIBRATION				
mV/V	Predicted Force Values	Read on Meter LB S/N 12140	Difference	%
-0.03999	47.92	47.93	-0.01	-0.016%
-0.07998	95.91	95.87	0.04	0.046%
-0.19995	239.88	239.68	0.20	0.083%
-0.39991	479.80	479.38	0.42	0.089%
-0.79979	959.51	958.74	0.77	0.080%
-1.19970	1439.13	1438.12	1.01	0.070%
-1.59962	1918.64	1917.51	1.13	0.059%
-1.99952	2398.04	2396.88	1.16	0.048%
-2.39942	2877.35	2876.25	1.10	0.038%
-3.19927	3835.81	3835.06	0.75	0.020%
-3.99901	4793.94	4793.81	0.13	0.003%
-4.39888	5272.96	5273.14	-0.18	-0.003%
Programmed @ 0.5000				

Figure 11 2-PT mV/V Values Entered into the 4215 Meter

The errors change quite a bit when one elects to use just a 2-pt span. We discussed this earlier, though here is another example where the values are better the closer one gets to capacity and deviate quite a bit throughout the range. Thus, I would argue that a 5-pt calibration is superior, though still significantly flawed compared with the coefficients in the formula for the calibration report.

Conclusion

Suppose the end goal is the best accuracy available. In that case, the recommendation will be a 4215 or HADI indicator, an ASTM E74 calibration, and software to convert mV/V values to Engineering units or a meter that allows coefficients to be entered. In these systems, we specify the accuracy from anywhere of 0.005 % to 0.025 % of full scale. These do not include drift effects, which is usually better than 0.02 % on these systems. For other systems that have a 5 or 10 pt. calibration and a meter is used to span the readings. We typically do not get better than 0.1 % of full scale if the calibration frequency is one year and have had several systems that can maintain 0.05 % of full scale on a six-month or less calibration interval. Taking a calibration report in mV/V and entering the mV/V values into the meter carries additional error that is very different to quantify based on the randomness of the points selected, and the error can vary. The actual results will vary on how much the system is used and on the individual components of the system.